

Certificate of Analysis

Company: Vermont Kind Manufacturing

Sample ID: MANU0027-C0054-1

Lot: N/A

Report Date: 11/10/2023

Matrix: Concentrate

Date Analyzed: 11/8/2023

Customer ID: 210614-02

Date Sampled: N/A

Analyst: 048

Grower License #: MANU0027

Date Received: 10/25/2023

Report ID: C231025AS

Residual Solvents Summary

Residual Solvent	LOQ (µg/g)	Results (µg/g)
Benzene	0.20	<LOQ
Chloroform	6.00	<LOQ
Methylene Chloride	500.00	<LOQ
Trichloroethylene	500.00	<LOQ
Acetone	40.00	<LOQ
Acetonitrile	500.00	<LOQ
Propane	500.00	<LOQ
Butane	500.00	<LOQ
Ethanol	500.00	735.96
Ethyl acetate	500.00	<LOQ
Ethyl Ether	500.00	<LOQ
Heptane	500.00	<LOQ
Hexane	30.00	<LOQ
Isopropyl Alcohol	500.00	<LOQ
Methanol	300.00	<LOQ
Pentane	500.00	<LOQ
Toluene	90.00	<LOQ
Total Xylenes	200.00	<LOQ

LOQ = The lowest quantity that this method can reliably detect. Any residual solvent that was not detected is assumed to be less than the stated LOQ (<LOQ).

Residual Solvent Methodology: Headspace Sampler, Gas Chromatography-Mass Spectrometry (GC-MS), using Perkin Elmer Clarus® SQ8 GC MS

Reagent Blanks: < LOQs for all analytes



This report shall not be reproduced except in full without approval of the laboratory. This is to provide assurance that parts of a report are not taken out of context. Results apply to the samples as received.

Certified by: *Luke E. M.*
 Luke Emerson Mason (Laboratory Director, Bia Diagnostics)